



**QPS Evaluation Services Inc.**  
**Testing, Certification and Field Evaluation Body**  
**Accredited in Canada, the USA, and Internationally**

**File**  
**QFC67241**

**CERTIFICATE OF COMPLIANCE**  
**(ISO TYPE 1 CERTIFICATION SYSTEM)**  
**FIELD CERTIFICATION**

Issued to	Boreal Laser Inc.
Address	12846-146 Street NW Edmonton, AB T5L 2H7
Project Number	QFC67241-1
Product	Gas Measurement Device with Optical Measurement Head Assembly
Model Number	GasFinder3-MC with OPX-ZZZZZ (ZZZZZ Is 5 Digit Serial Number), S/N 24656
Ratings	24VDC, 1A
Markings	Controller: Ex nA IIC T5 Tamb<65C Gc / AEx nA IIC T5 Gc Tamb<65C; Head: Ex ib IIC T5 Tamb<70C Gb / AEx ib IIC T5 Tamb<70C Gb
Applicable Standards	CSA C22.2 No. 60079-11-14 CSA C22.2 No. 60079-0-15 ISA-60079-11 (12.02.01)-2009 ANSI/ISA 60079-0 (12.00.01)-2013 ISA 12.12.01 2015 CSA C22.2 No. 61010-1-12-3 <sup>rd</sup> UL 508A 2nd Edition 2014
Inspection Location	Same as Applicant

**Statement of Compliance:** The product(s) identified in this Certificate and described in the Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Field Certification Label bearing the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.



**Certification Labels Used:**

This Certificate covers only the units bearing the following Certification Label number(s): **QD 24656**

**Issued By:** William Stephenson, Project Eng.

Signature:

Date: June 24, 2020